DIATOME

Diamond Knives

Cleaning procedure for our cryo diamond knives


Dear customer,

This procedure serves for the cleaning of our **cryo immuno** and **cryo 25° knives**.


We are at your disposal for any further assistance you might require.



Remove the knife stage from the cryo chamber, leave the knives in the stage, wash under tap water for warming up.



Bevel a polystyrene stick with a perfectly clean razer blade.



Clean the knife with 15% ethanol, keep wet.



Rince with 100% ethanol.



Dry with a dustblower or canned air.

Tel: +49 (0)89 18 93 668 0 • Fax: +49 (0)89 18 93 668 29 Mail: Info@ScienceServices.de • Web: ScienceServices.de